



Ref. No. VU/Tender/micro/18/25

Dated-20/02/2025

Tender Notice


Sealed quotations (mentioning the tender ref. no along with the last date of submission of the tender on the envelope) are invited from the reputed manufacturers, suppliers and vendors for the supply and installation of **Bacteriological Incubator (Digital Control) (300 Litter)** at Department of Microbiology, Vidyasagar University, Paschim Medinipur- 721102. The instrument will be purchased under the project grant of Department of Microbiology. The quotations are to be submitted only to the **HOD, Department of Microbiology, Vidyasagar University, Paschim Medinipur- 721102. Last date of submission of quotation is on 2.03.25**

Eligibility criterion of participation in the tender:

- a) Bonafide, resourceful and reliable Vendors.
- b) An undertaking should be given stating thereby that the firm has not been debarred or penalized for any reasons out of work by any Government Department.
- c) Subletting of suppliers is strictly prohibited.
- d) The prospective bidders must have valid upto date clearance of Income Tax return, Professional Tax clearance certificate, P. Tax (Deposit Challan), PAN Card, GST registration certificate. (Enclose latest copies).
- e) The suppliers who have been delisted or debarred by any government department shall not be eligible in any way.
- f) Past performance for vender will be judged (please attached three largest clients other than the Vidyasagar University).
- g) Rate is to be inclusive of all taxes and transport charges etc.
- h) The University reserves the right to accept or reject any quotation without assigning any reason whatsoever.
- i) Tender remains valid for six months.

Fax : (91) 03222 – 275329/275297/263206

Email : Vidya295@mail.vidyasagar.ac.in // Website : <http://www.vidyasagar.ac.in>


HEAD
Dept of Microbiology
VIDYASAGAR UNIVERSITY
Midnapore-721102, W.B.



VIDYASAGAR UNIVERSITY
MIDNAPORE WEST BENGAL PIN 721102

Details specifications are given below:
Statement showing the Specifications and quantity

Sl. No.	Name of the items	Specification	Quantity
01	Bacteriological Incubator (Digital Control) (300 Litter)	<ul style="list-style-type: none">• Operated on 220V/50Hz.• Inner size 24"(w) x24"(d) x32"(h).• 4nos shelves 300 litter.• Speed of shaker is variable from 30 RPM to 250 RPM. The tray of the shaker is also made of S.S. Sheet with Lotus type clamps to hold flasks. Unit is fitted with Digital RPM meter, 999 min digital time.• Double walled construction, inside wall made of thick Stainless Steel Sheet. Outside is made of mild steel sheet. Insulated with glass wool for good temperature stability and duly powder coating painted.• Heating coil, temperature will be controlled by a imported thermostat & double display digital pid controller with circulating system.• Temperature accuracy: $\pm 1^{\circ}\text{C}$, Temperature ambient +5 to 65°C. Inner a glass door for looking working sample at 37°C temperature.• The front heavy door with gasket should also highly insulated by glass wool to maintain temperature uniformly.• The system will be adjustable shelves, supplied with, indicating lamps, main cord plugs top & 4 nos. wheel etc.	1


HOD/PI

Department of Microbiology

HEAD
Dept of Microbiology
VIDYASAGAR UNIVERSITY
Midnapore-721102, W.B.

Fax : (91) 03222 – 275329/275297/263206

Email : Vidya295@mail.vidyasagar.ac.in // Website : <http://www.vidyasagar.ac.in>